



DEPARTMENT OF PHYSICS
MOHANLAL SUKHADIA UNIVERSITY
M.B. CAMPUS, UDAIPUR – 313 001(INDIA)

Email: physics@mlsu.ac.in

Ref.: Phy/CS/2025/102

Dated: 12/09/2025

Corrigendum to Open Tender Notice

Deadline extension for purchasing of Plasma Surface Treatment System and clarification

The **deadline** for submission of quotes for the limited tender notice vide number Phy/CS/2025/99 dated 06/09/2025, UBN: SUU2526GSOB00081, is hereby extended from September 15, 2025 2:00 pm to September 16, 2025 2:00 pm, will be opened on the same day i.e. September 16, 2025 at 3:00 pm.

Clarification on Specifications:

Specification(s)	Existing	Amended
Frequency	13.35 MHz	13.56 MHz
Metering valve	Should be able to qualitatively control the gas flow and chamber pressure.	Should be able to qualitatively control the gas flow and chamber pressure. OR Needle valve with equivalent performance to metering valve, but, metering valve will be preferred.
Chamber isolation and ventilation	Should have a 1/8th inch national pipe thread (NPT), a 1/8th inch NPT 3-way valve should be there for quick in bleeding of gas, isolating the chamber and ventilating.	Should have a 1/8th inch national pipe thread (NPT), a 1/8th inch NPT 3-way valve should be there for quick in bleeding of gas, isolating the chamber and ventilating OR Comprising of different venting valve with equivalent performance.

Rest of the Terms and Conditions are remained same.

8/12/31/2025
HEAD

Department of Physics

विभागाध्यक्ष
भौतिक शास्त्र विभाग
मोहनलाल सुखाड़िया विश्वविद्यालय
उदयपुर (राज.)